

出口销售证明

Certificate of Export

中华人民共和国
PEOPLE'S REPUBLIC OF CHINA
医疗器械产品出口销售证明
CERTIFICATE FOR EXPORTATION OF MEDICAL
PRODUCTS

证书编号: 湘药监械出 20200064 号
Certificate NO.: HNMPA 20200064

产品名称: 医用外科口罩
Product(s): Medical Surgical Face Mask

规格型号: 平面形 (耳挂式、绑带式), 尺寸规格为 17cm × 9cm, 允差 5%
Model: Flat form (ear loop and lexpert-up) Size, 17cm × 9cm, deviation 5%

产品注册或备案凭证号: 湘械注准 20202140333
Registration certificate(s): HNMPA Certified 20202140333

生产企业: 湖南豌豆医疗用品有限公司
Manufacturer: Hunan Wondo Medical Supplies Co., Ltd.

生产企业住所: 长沙经济技术开发区人民东路二段 169 号先进储能节能创意示范
产业园 12 栋 502
Address of manufacturer: Building 12th, Advanced Energy Storage and Energy
Saving of Creative Industrial Park, No. 169, 2nd, Area East Renmin Road,
Changsha Economic and Technological Development Zone, Hunan,
Province, China.

生产许可或备案凭证号: 湘食药监械生产许 20190009 号
Manufacturing License(s): Hunan CFDA Production Permit No. 20190009 号

兹证明上述产品已准许在中国生产和销售。
This is to certify that the above products have been registered
to be manufactured and sold in China.

证明有效期至: 2022 年 4 月 12 日
This certification valid until: Apr. 12, 2022



CE 认证流程说明

CE Certificate Procedure

豌豆医疗 CE 认证流程


- 1、自我符合性声明（包括欧代信息、产品类别、引用标准、符合欧盟法规 MDR 等信息）；
- 2、由欧盟委员会授权的（中国仅 3 家）、具有 MDR 审核资质的公告机构南德认证检测（TÜV SÜD）出具产品的相关检测报告；
- 3、欧代完成审定并递交德国主管当局（药监局）备案，德国药主管当局出具官方签字版的备案凭证，该凭证代表我公司产品符合欧盟标准并具有 CE 出口资质。

湖南豌豆医疗用品有限公司

2020 年 3 月 25 日

EC 符合性声明

Declaration of Conformity

 European Commission			
Declaration of Conformity			
Manufacturer: Hunan Wondo Medical Supplies Co., Ltd.			
Address: Building 12th, Advanced Energy Storage and Energy Saving of Creative Industrial Park, No. 169, 2nd, Area East Renmin Road, Changsha Economic and Technological Development Zone, Hunan, Province, China			
EU Authorised Representative:			
Osmunda Medical Technology Service GmbH			
Address: Von Oppen-Weg 15, 14476 Potsdam, Germany			
DIMDI code: DE/0000047267			
Device: Medical surgical face mask			
Model: Flat form (ear loop and lexpert-up); Folded form			
Classification (MDR, Annex VIII): Class I			
Conformity assessment route: ANNEX II+ANNEX III			
We herewith declare that the above mentioned product meet the provisions of the following regulation (EU) MDR 2017/745. All supporting documentations are retained under the premises of the manufacturer.			
			
General applicable directives:			
Medical Device Directive: REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on medical devices.			
Standard Applied:			
EN ISO 13485:2016	EN ISO 14971:2012	EN 14683:2019 +AC: 2019	BS EN ISO 15223-1:2016
EN ISO 10993-1:2009/AC:2010	EN ISO 10993-5:2009	ISO 10993-10:2010	EN 1041:2008
Place, Date of Issue:		Signature:	
Changsha, 30 th March 2020			
		Manager representative	

TÜV 检测报告

TÜV Test Report

Test Report No.: 721653785-11
Report Date: 23 April 2020



SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME Hunan Wondo Medical Supplies Co., Ltd

CLIENT ADDRESS Building 12th, Advanced Energy Storage and Energy Saving of Creative Industrial Park, No. 169, 2nd, Area East Renmin Road, Changsha Economic and Technological Development Zone, Hunan, Province, China.

TEST PERIOD 08-Apr-2020~16-Apr-2020

Prepared By

Bella Xu

(Bella Xu)
Report Drafter

Authorized By



(Leo Liu)
Authorized Signatory

Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested. (3) The test report shall not be reproduced except in full without the written approval of the laboratory. (4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food_chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

TÜV®

Page 1 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



TEST REPORT

Sample Description : Medical Surgical Face Mask
Sample Quantity : 50 pieces
Lot Number/Batch Code : 20200324
Specification : Flat form ear loop
Size : 17cm*9cm
Type of Mask : Type IIR
Brand Name : /

Remark: The above information was provided by applicant.

Summary of Test Results

No.	Test Item	Test Standard	Judgement
1	Bacterial Filtration Efficiency (BFE) Test	EN 14683:2019+AC:2019(E) Annex B	Pass
2	Differential Pressure Test	EN 14683:2019+AC:2019(E) Annex C	Pass
3	Synthetic Blood Penetration Test	ISO 22609:2004	Pass
4	Microbial Cleanliness Test	EN 14683:2019+AC:2019(E) Annex D	Pass

Note: Pass = Meet customer requirements;

Fail = Fail customer requirements;

= No comment;

N.D. = Not detected.

Photo of Samples



Chemical/Microbiology Laboratory:
TUV SUD Products Testing (Shanghai) Co.,
Ltd.
B-34, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SUD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200070 P.R.China



Page 2 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



Results

No.	Test Item	Test Result
1	Bacterial Filtration Efficiency (BFE) Test	Specimen 1#: 99.5% Specimen 2#: 99.4% Specimen 3#: 99.4% Specimen 4#: 99.5% Specimen 5#: 99.5%
2	Differential Pressure Test	26.5 Pa/cm ²
3	Synthetic Blood Penetration Test	Specimen 1#~13#: None seen
4	Microbial Cleanliness Test	Specimen 1#: 5 CFU/g Specimen 2#: 27 CFU/g Specimen 3#: 13 CFU/g Specimen 4#: 10 CFU/g Specimen 5#: 8 CFU/g

Bacterial Filtration Efficiency (BFE) Test

1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of mask.

2. Sample description was given by client

Sample description : Medical Surgical Face Mask
Specification : Flat form ear loop
Lot Number : 20200324
Sample Receiving Date : 2020-04-08

3. Test Method

EN 14683:2019+AC:2019(E) Annex B

4. Apparatus and materials

- 4.1 *Staphylococcus aureus* ATCC 6538.
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.

5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.

Chemical/Microbiology Laboratory:
TUV SUD Products Testing (Shanghai) Co.,
Ltd.
8-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SUD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China



Page 3 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



6. Procedure

- 6.1 Preparation of the bacterial challenge: Dilute the culture in peptone water to achieve a concentration of approximately 5×10^5 CFU/mL.
- 6.2 Adjust the flow rate through the Anderson sampler to 28.3 L/min.
- 6.3 Deliver the challenge to the nebulizer using a syringe pump. Purge tubing and nebulizer of air bubbles.
- 6.4 Perform a positive control run without a test specimen to determine the number of viable aerosol particles being generated. The mean particle size (MPS) of the aerosol will also be calculated from the results of these positive control plates.
 - 6.4.1 Initiate the aerosol challenge by turning on the air pressure and pump connected to the nebulizer. Immediately begin sampling the aerosol using the Anderson sampler.
 - 6.4.2 Time the challenge suspension to be delivered to the nebulizer for 1 min.
 - 6.4.3 Time the air pressure and Anderson sampler to run for 2 min.
 - 6.4.4 At the conclusion of the positive control run, remove plates from the Anderson sampler.
- 6.5 Place new agar plates into Anderson sampler and clamp the test specimen into the top of the Anderson sampler, with the inside of the specimen facing towards the bacterial challenge (test area: 77cm^2).
- 6.6 Repeat the challenge procedure for each test specimen.
- 6.7 Repeat a positive control after completion of the sample set.
- 6.8 Perform a negative control run by collecting a 2 min sample of air from the aerosol chamber. No bacterial challenge should be pumped into the nebulizer during the collection of the negative control.
- 6.9 Incubate agar plates at $(37 \pm 2)^\circ\text{C}$ for (20 to 52) h.
- 6.10 Count each of the six-stage plates of the Anderson sampler.

7. Calculation

Total the count from each of the six plates for the test specimens and positive controls, as specified by the manufacture of Anderson sampler. The filtration efficiency percentages are calculated as follows:

$$\text{BFE} = (C - T) / C \times 100$$

T is the total plate count for the test specimen.

C is the mean of the total plate counts for the two positive controls.

Chemical/Microbiology Laboratory:
TUV SUD Products Testing (Shanghai) Co.,
Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201105
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SUD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China



Page 4 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



8. Test results*

Stage Number	P Value	Positive Control (A)	Positive Control (B)	Negative Control	Specimen 1#	Specimen 2#	Specimen 3#	Specimen 4#	Specimen 5#
1		33	55	0	0	0	0	0	0
2		120	246	0	2	2	2	2	2
3		112	294	0	1	0	1	1	1
4		119	298	0	0	0	0	0	0
5		1341	1219	0	4	4	3	4	4
6		347	450	0	4	7	7	5	4
Total (T), CFU		2072	2562	<1	11	13	13	12	11
Average (C), CFU		$2.3 \times 10^2 = (P_A + P_B) / 2$							
BFE, %		99.5 99.4 99.4 99.5 99.5							
Requirements		≥ 98							
Remarks		P is the value of corresponding corrected particle counts as specified by the manufacturer of the cascade impactor. T is the total of P value for the test specimen. C is the mean of the total of P value of the two positive controls.							

Chemical/Microbiology Laboratory:
TUV SÜD Products Testing (Shanghai) Co., Ltd.
8-3/4, No. 1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SÜD Certification and Testing (China) Co., Ltd.
No. 151 Heng Tong Road Shanghai
200070 P.R. China



Page 5 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



Differential pressure Test

1. Purpose

The purpose of the test was to measure the differential pressure of masks.

2. Sample description was given by client

Sample description : Medical Surgical Face Mask
Specification : Flat form ear loop
Lot Number : 20200324
Sample Receiving Date : 2020-04-08

3. Test Method

EN 14683:2019+AC:2019(E) Annex C

4. Apparatus and materials

Differential pressure testing instrument

5. Test specimen

- 5.1 Test specimen are complete masks or shall be cut from masks. Each specimen shall be able to provide 5 different circular test areas of 2.5 cm in diameter.
5.2 Prior to testing, condition all test specimens for a minimum of 4 h at $(21 \pm 5)^\circ\text{C}$ and $(85 \pm 5)\%$ relative humidity.

6. Procedure

- 6.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.
6.2 The pretreated specimen is placed across the orifice (total area 4.9cm^2 , test area diameter 25mm) and clamped into place so as to minimize air leaks.
6.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.
6.4 The differential pressure is read directly.
6.5 The procedure described in steps 6.1-6.4 is carried out on 5 different areas of the mask and readings averaged.

Results:

Specimen	Test Results* (Pa/cm ²)	Average (Pa/cm ²)	Requirements	Judgement
1#	27.1	26.5	< 60	Pass
2#	26.6			
3#	24.3			
4#	25.0			
5#	29.5			

Chemical/Microbiology Laboratory:
TUV SUD Products Testing (Shanghai) Co.,
Ltd.
8-3/4, No.1599 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SUD Certification and Testing
(China) Co., Ltd.
No.151 Hong Tong Road Shanghai
200 070 P.R.China



Page 6 of 11

TUV
SUD
1
*

Test Report No.: 721653785-11
Report Date: 23 April 2020



Synthetic Blood Penetration Test

1. Purpose

For evaluation of resistance of masks to penetration by a fixed volume of synthetic blood at a high velocity.

2. Sample description was given by client

Sample description : Medical Surgical Face Mask
Specification : Flat form ear loop
Lot Number : 20200324
Sample Receiving Date : 2020-04-08

3. Test Method

ISO 22609:2004

4. Apparatus and materials

- 4.1 Synthetic blood.
- 4.2 Tensiometer.
- 4.3 Synthetic blood penetration test apparatus;
- 4.4 Targeting plate.
- 4.5 Air pressure source.
- 4.6 Ruler.
- 4.7 Balance.
- 4.8 Controlled temperature and humidity chamber.

5. Test specimen

- 5.1 As requested by client, take a total of 13 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4h at $(21 \pm 5)^{\circ}\text{C}$ and $(85 \pm 5)\%$ relative humidity.

6. Procedure

- 6.1 Prepare the synthetic blood (40~44 mN/m) for the test.
- 6.2 Determine the density of the synthetic blood.
- 6.3 Fill the reservoir with new synthetic blood.
- 6.4 Position the test specimen 30.5 cm (12 in.) from the exit of the canula.
- 6.5 Set the reservoir pressure to the approximate pressure.
- 6.6 Place the targeting plate approximately 1 cm away from the mask.
- 6.7 Set the valve timer to 0.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).
- 6.8 Set the valve timer to 1.5 s. Collect and weigh the amount of fluid delivered (before the targeting

Chemical/Microbiology Laboratory:
TUV SUD Products Testing (Shanghai) Co.,
Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SUD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200070 P.R.China



Page 7 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



hole).

- 6.9 Calculate the difference in weight of the two spurts. For a test fluid with a density of 1.003, Table 1 gives the target difference in weight plus lower and upper limits for a velocity range within 2% of the target.

Table 1 Target weight difference

Fluid Pressure (mmHg)	Weight difference for 1s difference in spurt duration (g)		
	Min.	Target	Max.
120	3.002	3.063	3.124

- 6.10 Adjust the reservoir pressure and repeat steps 6.7 to 6.9 until the weight difference is within the target range.
- 6.11 Record the weight difference for the spurts exiting the nozzle.
- 6.12 Record the pressure in the reservoir.
- 6.13 Set the valve time to 0.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.14 Set the valve time to 1.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.15 The difference in weight between the 0.5 s and 1.5 s spurts through the targeting plate shall be within +2 % ~ -5 % of the difference in weight from the nozzle.
- 6.16 If the differential weight is less than 95 % of the weight difference exiting the nozzle, check the aim of the stream to make sure it is passing cleanly through the targeting hole.
- 6.17 If the differential weight is more than 102 % of the weight difference exiting the nozzle, repeat the weight measurements exiting the nozzle (steps 6.7 to 6.11).
- 6.18 For standard synthetic blood, the timer duration can be estimated using the formula:
(p is the density of the test fluid.) $t = 0.5 + (2 \times p - g \text{ at } 0.5 \text{ s}) / (g \text{ at } 1.5 \text{ s} - g \text{ at } 0.5 \text{ s})$.
- 6.19 Record the timer setting to use as the starting point for subsequent testing.
- 6.20 Mount a test specimen on the specimen holding fixture. If the mask contains pleats, spread the pleats out when mounting the mask onto the fixture to present a single layer of material as the target area.
- 6.21 Squirt the synthetic blood onto the test specimen for the calculated time. Ensure that the synthetic blood hits the target area of mask.
- 6.22 Inspect the inside surface for synthetic blood penetration within 10 s of squirting the synthetic blood against the target area.
- 6.23 Report the results (none / penetration) for each test specimen at the test pressure.

Chemical/Microbiology Laboratory:
TUV SUD Products Testing (Shanghai) Co.,
Ltd.
8-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SUD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China



Page 8 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



Results:

Specimen	Test Results*	Requirements	Judgement
1#	None Seen	Pass Pressure at 16.0 kPa (120mmHg)	Pass
2#	None Seen		Pass
3#	None Seen		Pass
4#	None Seen		Pass
5#	None Seen		Pass
6#	None Seen		Pass
7#	None Seen		Pass
8#	None Seen		Pass
9#	None Seen		Pass
10#	None Seen		Pass
11#	None Seen		Pass
12#	None Seen		Pass
13#	None Seen		Pass



LANG
BY
UT

Chemical/Microbiology Laboratory:
TUV SUD Products Testing (Shanghai) Co.,
Ltd.
8-34, No.1509 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SUD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China



Page 9 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



Microbial Cleanliness Test

1. Purpose

The purpose of the test was to measure microbial cleanliness of mask.

2. Sample description was given by client

Sample description : Medical Surgical Face Mask
Specification : Flat form ear loop
Lot Number : 20200324
Sample Receiving Date : 2020-04-08

3. Test Method

According to EN ISO 11737-1:2018 to determine the microbial cleanliness of mask material, and refer to the procedure as described in EN 14683:2019+AC:2019(E) Annex D

4. Apparatus and materials

- 4.1 Orbital shaker.
- 4.2 0.45 um filter.
- 4.3 Tryptic Soy Agar (TSA).
- 4.4 Sabouraud Dextrose Agar (SDA) with chloramphenicol.
- 4.5 Formula of Extraction Liquid: 1g/L peptone, 5g/L NaCl and 2g/L Tween 20.
- 4.6 Extraction apparatus.

5. Test specimen

- 5.1 As requested by client, take a total of 5 mask samples.
- 5.2 Mask samples for testing are provided in the original primary packaging.
- 5.3 Condition at (18 to 26)°C and (45 to 65)% relative humidity during testing.

6. Procedure

- 6.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.
- 6.2 The mask is aseptically removed from the packaging and placed in a sterile 500 mL bottle containing 300 mL of extraction liquid.
- 6.3 The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm.
- 6.4 After extracting, 100mL of the extraction liquid is filtered through a 0.45 um filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungi enumeration.
- 6.5 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.
- 6.6 Calculate the colonies of each agar plate.

7. Calculation

For each test specimen calculate the microbial cleanliness as follows by counting the total colonies of the TSA and SDA plates.

Chemical/Microbiology Laboratory:
TUV SÜD Products Testing (Shanghai) Co.,
Ltd.
8-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TUV SÜD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China



Page 10 of 11

Test Report No.: 721653785-11
Report Date: 23 April 2020



Results*:

Specimen	Colonies of the TSA Plate	Colonies of the SDA Plate	Microbial Cleanliness, (CFU/g)	Requirements	Judgement
1#	4	1	5	According to EN ISO 11737-1:2018 the microbial cleanliness of the mask shall be ≤ 30 CFU/g tested.	Pass
2#	15	12	27		
3#	7	6	13		
4#	8	2	10		
5#	7	1	8		

Note:

- 1.*denotes this test was carried out by external laboratory assessed as competent.
- 2.This report is for internal use only such as internal scientific research ,education, quality control, product R&D.

-END OF THE TEST REPORT-



Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing (China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China



Page 11 of 11

CE 备案凭证

CE Filing Certificate

Anlage 1
(zu § 4 Abs. 1 Nr. 1 DIMDIV)
Formularnummer 00301438

Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG

Formblatt für Medizinprodukte, außer In-vitro-Diagnostika
Form for Medical Devices except In Vitro Diagnostic Medical Devices

Zuständige Behörde / Competent authority	
Code DE/CA76	
Bezeichnung / Name Landesamt für Arbeitsschutz, Verbraucherschutz und Gesundheit, Abteilung Gesundheit, Dezernat G4	
Staat / State Deutschland	Land / Federal state Brandenburg
Ort / City Zossen	Postleitzahl / Postal code 15806
Straße, Haus-Nr. / Street, house no. Wünsdorfer Platz 3	
Telefon / Phone +49-331-8683852	Telefax / Fax +49-331-8683865
E-Mail / E-mail medizinprodukte@lavg.brandenburg.de	
Anzeige / Notification	
Registrierdatum bei der zuständigen Behörde Registration date at competent authority	Registriernummer / Registration number
Typ der Anzeige / Notification type ☒ Erstanzeige / Initial notification ☐ Änderungsanzeige / Notification of change ☐ Widerrufsanzeige / Notification of withdrawal	
Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn	
Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG ☐ Hersteller / Manufacturer ☒ Bevollmächtigter / Authorised Representative ☐ Einführer / Importer ☐ Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG ☐ Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV ☐ Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG	

Anzeigender / Reporting organisation (person)	
Code	DE/0000047267
Bezeichnung / Name	Osmunda Medical Technology Service GmbH
Staat / State	Deutschland
Land / Federal state	Brandenburg
Ort / City	Potsdam
Postleitzahl / Postal code	14476
Straße, Haus-Nr. / Street, house no. Von Oppen-Weg, 15	
Telefon / Phone	030 50590627
Telefax / Fax	
E-Mail / E-mail	eu@osmundacn.com

Hersteller / Manufacturer	
Bezeichnung / Name	Hunan Wondo Medical Supplies Co., Ltd
Staat / State	CN
Ort / City	Changsha, Hunan
Postleitzahl / Postal code	410000
Straße, Haus-Nr. / Street, house no. Building 12th, Advanced Energy Storage and Energy Saving of Creative Industrial Park, No. 169, 2nd, Area East Renmin Road, Changsha Economic and Techn	
Telefon / Phone	0086-731-86910566
Telefax / Fax	
E-Mail / E-mail	

Sicherheitsbeauftragter für Medizinprodukte nach § 30 Abs. 2 MPG 9) Safety officer for medical devices pursuant to § 30 (2) Medical Devices Act, MPG	
Bezeichnung / Name	Min Yang
Staat / State	Deutschland
Land / Federal state	Berlin
Ort / City	Berlin
Postleitzahl / Postal code	10787
Straße, Haus-Nr. / Street, house no. Keithstr. 2-4	
Telefon / Phone	030 5059 0627
Telefax / Fax	
E-Mail / E-mail	min.yang@osmundacn.com

Anlage 1
(zu § 4 Abs. 1 Nr. 1 DIMDIV)
Formularnummer 00301438

Vertreter / Deputy (optional)	
Bezeichnung / Name	
Telefon / Phone	Telefax / Fax
E-Mail / E-mail	
☐ Erstanzeige / Initial notification ☒ Änderungsanzeige / Notification of change	

Medizinprodukt (Erstmaliges Inverkehrbringen) / Medical device (First placing on the market)	
Klasse / Class	☒ I ☐ I - steril / sterile ☐ I - mit Messfunktion / with measuring function ☐ I - steril und mit Messfunktion / sterile and with measuring function ☐ IIa ☐ IIb ☐ III ☐ III - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012 ☐ Aktives implantierbares Medizinprodukt / Active implantable medical device ☐ Aktives implantierbares Medizinprodukt - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 Active implantable medical device - manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012
App (Software auf mobilen Endgeräten)	☐ ja / yes ☒ nein / no
Nummer(n) der Bescheinigung(en) / Certificate number(s)	
Handelsname des Produktes / Trade name of the device	Medical surgical face mask
Produktbezeichnung / Name of device	
Nomenklaturcode / Nomenclature code	12-458
Nomenklaturbezeichnung / Nomenclature term	Maske, Chirurgie
Kategoriecode / Category code	10
Kategorie / Category	Produkte zum Einmalgebrauch
Kurzbeschreibung deutsch / German short description	
Kurzbeschreibung englisch / English short description	The medical surgical face mask is suitable for medical workers basic protection and related workers under environment with the risk of body fluid and spatter.

Medizinprodukte (Aufbereiten) / Medical devices (Reprocessing)	
☐	Semikritische Medizinprodukte / Semicritical medical devices ☐ Gruppe A / Group A ☐ Gruppe B / Group B
☐	Kritische Medizinprodukte / Critical medical devices ☐ Gruppe A / Group A ☐ Gruppe B / Group B ☐ Gruppe C / Group C Nummer der Bescheinigung / Certificate number
☐	Sterilisationsverfahren / Sterilisation procedures ☐ Dampfsterilisation / Steam sterilisation ☐ Gassterilisation / Gas sterilisation ☐ Strahlensterilisation / Radiation sterilisation ☐ andere / others Angewandtes Verfahren / Applied procedure

Ich versichere, dass die Angaben nach bestem Wissen und Gewissen gemacht wurden.
I affirm that the information given above is correct to the best of my knowledge.

Ort
City

Berlin

Datum
Date

2020-04-05

Name

Min Yang

Unterschrift
Signature

Bearbeitungsvermerke / Processing notes	
Nur von der zuständigen Behörde auszufüllen / To be filled in only by the competent authority	
Bearbeiter / Person responsible	Telefon / Phone

包装信息

Packaging Information



Bag: 18cmx29cm
50 PCs/Bag



Box: 8cmx11cmx20cm
1 Bag/Box



QTY/Box: 50Pcs

QTY of Boxes/Carton: 40 Boxes

QTY/Carton: 2000 PCs

G.W: 8.7KG

N.W: 6.4KG

Measure: 57cmx42cmx34cm